



To: DC Managers
From: Sam Buckley
Date: April 3, 2017
RC: 17-069A

URGENT!!! UPDATED EXPIRATION DATES ON LOTS!!! URGENT!!!

FDA/SUPPLIER CLASS OF RECALL: Not Yet Classified
LEVEL OF NOTIFICATION: Consumer
SUPPLIER: Mylan - #12548

Description	Lot #	Exp Date	NDC	UPC	Econo #
EPIPEN 0.3MG 0.3ML 1:1M 2PK	6GM199 10/31/17; 6GM198 10/31/17; 6GM091 10/31/17; 6GM088 10/31/17; 6GM087 10/31/17; 6GM082 09/30/17; 6GM081 09/30/17; 6GM072 09/30/17; 5GM640 05/31/17; 5GM631 04/30/17;		49502050002	34950250002	1412022
EPIPEN JR .15MG .3ML 1:2M 2PK	6GN215	09/30/17	49502050102	34950250102	1411107
	5GN773	04/30/17			
	5GN767	04/30/17			

Mylan is voluntarily recalling the above items/lots due to the result of two previously disclosed reports outside of the U.S. of failure to activate the device due to a potential defect in a supplier component. This recall is to the Consumer level. Affected product started shipping December 2015.

Distribution Centers should immediately check inventory, quarantine, and discontinue distribution of the affected product. Complete and return the Business Reply Card form to Stericycle. Please complete the BRC even if you do not have product. Return the product to Stericycle using the prepaid shipping label. If you have affected product and need a return kit, contact Stericycle at 877-650-3494.

Stericycle
 Attn: Event # 4403
 2670 Executive Drive, Suite A
 Indianapolis, IN 46241
 DEA # RS0331607

McKesson Customer, please examine your inventory for the affected item completely and stop dispensing immediately. If you have the affected lot number(s), please contact Stericycle at 877-650-3494 to request a returns kit.

PLEASE NOTE: Customers currently participating in a McKesson administered Return to Vendor program should return this product to their designated returns processor. All others should follow the instructions provided by the manufacturer.

This recall is being conducted with the knowledge of the FDA.

Information contained in this document was provided by Mylan.

FOR McKESSON USE ONLY:

- Please follow Recall Guidelines as outlined in the Reclamation SOP for a CLASS I RECALL.
- PROCESS TO THE CONSUMER LEVEL.